



Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

Packaged Product: 1770			
Chloride Standard 0.0141 Normal (500ppm Cl) NIST Traceable			
Test	Target/UOM	Range	Result
PREPARED TO FORMULATION ON FILE	YES		YES
Sample ID	YES		
Tested 08/24/23			
Tester MHUS			
Recorded 08/24/23			
APPEARANCE AND COLOR	CLEAR LIQUID APPEARANCE		PASS
Sample ID			
Tested 08/24/23			
Tester MHUS			
Recorded 08/24/23			
Actual Density	0.99885 GM/ML	0.99862 - 0.99907	0.99895
Sample ID			
Tested 08/24/23			
Tester MHUS			
Recorded 08/24/23			
INSTRUMENTS USED DURING PREPARATION	INSTRUMENT		M91,VF2-A8,M55,BIDAT,PIDP1 18
Sample ID			
Tested 08/24/23			
Tester MHUS			
Recorded 08/24/23			
NORMALITY TO 4 DEC PLACES	0.0141 EQUIV/LITER	0.0138 - 0.0144	0.0140
Sample ID			
Tested 08/24/23			
Tester MHUS			
Recorded 08/24/23			
QCSP68:STANDARDIZATION SODIUM CHLORIDE N	YES QCSP68		YES
Sample ID			
Tested 08/24/23			
Tester MHUS			
Recorded 08/24/23			
CALCULATED PPM CHLORIDE (CL)	500.0 PP	489.0 - 511.0	496.3
Sample ID			
Tested 08/24/23			
Tester MHUS			
Recorded 08/24/23			
EXPIRATION DATE	MM/DD/YY MM/DD/YYYY		08/30/24
Sample ID			
Tested 08/24/23			
Tester MHUS			
Recorded 08/24/23			



Aqua Solutions, Inc.
 Ph: 281-479-2569
 Fax: 281-479-2790
 P.O. Box 70
 Deer Park, TX 77536
 Website: www.aquasolutions.org

Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

SHELF LIFE OF PRODUCT

1 YEAR
 TIME

1 YEAR

Sample ID

Tested 08/24/23

Tester MHUS

Recorded 08/24/23

Lot # 3082327

Made 08/24/23

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

Quality Manager