



Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

Packaged Product: LY160F POTASSIUM STANDARD 0.1 PPM IN METHANOL			
Test	Target/UOM	Range	Result
PREPARED TO FORMULATION			
	YES		YES
	YES		
Sample ID			
Tested	08/23/23		
Tester	AFRE		
Recorded	08/23/23		
CALCULATED PPM POTASSIUM			
	0.10	0.09 - 0.11	0.10
	MG/L		
Sample ID			
Tested	08/23/23		
Tester	AFRE		
Recorded	08/23/23		
APPEARANCE AND COLOR			
	CLEAR WATER WHITE		PASS
	APPEARANCE		
Sample ID			
Tested	08/23/23		
Tester	AFRE		
Recorded	08/23/23		
INSTRUMENTS USED DURING PREPARATION			
	LAB EQUIPMENT		P137,VF1-Z14
	INSTRUMENT		
Sample ID			
Tested	08/23/23		
Tester	AFRE		
Recorded	08/23/23		
EXPIRATION DATE			
	MM/DD/YYYY		09/28/23
	MM/DD/YYYY		
Sample ID			
Tested	08/23/23		
Tester	AFRE		
Recorded	08/23/23		
Lot # 3082246			
Made 08/23/23			

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.