

Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results

Customer PO # :

Packaged Product:	6937			
·	CONDUCT	IVITY STANDARD 2.0 uMHOS	**Make fresh	
Test		Target/UOM	Range	Result
PREPARED TO FORMULA	TION ON	YES		YES
FILE		YES		
Sample ID				
	3/21/23			
	PHI			
	3/21/23			
CONDUCTIVITY		2.0	1.9 - 2.1	2.0
		UMHOS		
Sample ID				
Tested 08	3/21/23			
	PHI			
	8/21/23			
QCSP30:CONDUCTIVITY		YES		YES
STD&MEASUREMENT		QCSP30		
Sample ID				
Tested 08	3/21/23			
Tester BI	PHI			
	3/21/23			
TRACEABLE TO N.I.S.T. SP	RM 3191	YES		YES
		YES		
Sample ID				
Tested 08	8/21/23			
Tester Bl	PHI			
Recorded 08	3/21/23			
APPEARANCE AND COLO	R	CLEAR WATER WHITE		PASS
		LIQUID		
		APPEARANCE		
Sample ID				
Tested 08	3/21/23			
	PHI			
Recorded 08	3/21/23			
INSTRUMENTS USED DUR	RING	INSTRUMENT		M-92,T-234
PREPARATION				
Sample ID				
	3/21/23			
	PHI			
	3/21/23			
EXPIRATION DATE		MM/DD/YY		11/21/23
		MM/DD/YYYY		
Sample ID				
	3/21/23			
Tester BI	PHI			
Recorded 08	3/21/23			
SHELF LIFE OF PRODUCT		3 MONTHS		3 MONTHS
		TIME		
Sample ID				
	3/21/23			
	PHI			
Recorded 08	3/21/23			



Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results

Customer PO # :

Lot #	3081856
Made	08/21/23

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

chard Stol

Quality Manager