

Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results

Customer PO # :

Packaged Product	. CIT011	F		
	•	CHLORINE STANDARD 175 B	PPM (CHLORAMINE T)	
Test		Target/UOM	Range	Result
PREPARED TO FORMU		YES	nange	YES
FILE		YES		120
Sample ID		120		
Tested	08/25/23			
Tester	CCOA			
Recorded	08/25/23			
PPM CHLORINE (CL2)	00/23/23	175.00	170.00 - 180.00	174.45
		PPM	170.00 100.00	174.45
Comete ID				
Sample ID	00/05/00			
Tested	08/25/23			
Tester Recorded	CCOA			
Recorded	08/25/23	CLEAR LIQUID		PASS
AFPEARANCE AND CO				PA33
		APPEARANCE		
Sample ID				
Tested	08/25/23			
Tester	CCOA			
Recorded	08/25/23			
INSTRUMENTS USED DURING		INSTRUMENT		M91,T-157,M55,BID00488,PID
PREPARATION				P40-58
Sample ID				
Tested	08/25/23			
Tester	CCOA			
Recorded	08/25/23			
Actual Density		0.99853 GM/ML	0.99796 - 0.99911	0.99857
Operate JD				
Sample ID	00/05/00			
Tested	08/25/23			
Tester	CCOA			
Recorded	08/25/23			
Lot # 30823				
Made 08/25	5/23			

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

huland SPA