Page: 1

Date: 08/22/23 at 1:09 PM

Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results

Customer PO #:

Packaged Product:	TEN098F		
	FLUORIDE STANDARD 2 PPM		
Test	Target/UOM	Range	Result
PREPARED TO FORMULA			YES
FILE	YES		
Sample ID			
Tested 0	8/22/23		
Tester M	MAT		
Recorded 0	8/22/23		
APPEARANCE AND COLO	OR CLEAR LIQUID		PASS
	APPEARANCE		
Sample ID			
Tested 0	8/22/23		
Tester M	MAT		
Recorded 0	8/22/23		
INSTRUMENTS USED DU	RING INSTRUMENT		P138,VF500-X43
PREPARATION			
Sample ID			
Tested 0	8/22/23		
Tester M	MAT		
Recorded 0	8/22/23		
EXPIRATION DATE	MM/DD/YYYY		09/27/23
	MM/DD/YYYY		
Sample ID			
Tested 0	8/22/23		
Tester M	MAT		
Recorded 0	8/22/23		
Lot # 3082251			
Made 08/22/2	23		

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

Quality Manager

hichard Stol