

Page: 1

Date: 08/23/23 at 4:25 PM

Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569

Customer PO #:

Customer PO #:				
Packaged Product:	OXY007F			
	LEVEL 2	CELL LIQUOR STANDARD		
Test		Target/UOM	Range	Result
PREPARED TO FORMULA	ATION ON	YES		YES
FILE		YES		
Sample ID				
	08/23/23			
	AFRE			
	08/23/23			
CALCULATED GPL SODI	UM	125.00	124.98 - 125.02	125.00
CHLORIDE		GPL		
Sample ID				
Tested (08/23/23			
Tester	AFRE			
	08/23/23			
CALCULATED GPL SODI	UM	180.0	179.9 - 180.1	180.0
HYDROXIDE		GPL		
Sample ID				
	08/23/23			
Tester /	AFRE			
Recorded (08/23/23			
CALCULATED GPL SODI	UM	0.3000	0.2995 - 0.3005	0.3000
CHLORATE		GPL		
Sample ID				
Tested (08/23/23			
Tester /	AFRE			
Recorded (08/23/23			
APPEARANCE AND COLOR		CLEAR LIQUID		PASS
		APPEARANCE		
Sample ID				
Tested (08/23/23			
Tester	AFRE			
Recorded (08/23/23			
INSTRUMENTS USED DU	IRING	INSTRUMENT		M100,VF2-A4
PREPARATION				
Sample ID				
	08/23/23			
	AFRE			
	08/23/23			
Actual Density		1.23794	1.23605 - 1.23982	1.23850
		GM/ML		
Sample ID				
	08/23/23			
	AFRE			
	08/23/23			
Lot # 308225	4			
Made 08/23/				



Page: 2

Date: 08/23/23 at 4:25 PM

Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO #:

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

Quality Manager