

Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

Packaged Product: VUL380F			
SODIUM HYDROXIDE 0.001 NORMAL IN METHANOL			
Test	Target/UOM	Range	Result
PREPARED TO FORMULATION ON	YES		YES
FILE	YES		
Sample ID			
Tested 08/24/23			
Tester CCOA			
Recorded 08/24/23			
SAMPLE COLLECTION & HANDLING	YES		YES
PER LAB040	YES		
Sample ID			
Tested 08/24/23			
Tester CCOA			
Recorded 08/24/23			
DATE OF SAMPLING FOR TEST	MM/DD/YYYY		08/24/23
PURPOSES	MM/DD/YYYY		
Sample ID			
Tested 08/24/23			
Tester CCOA			
Recorded 08/24/23			
AMBIENT TEMPERATURE DURING	25.0	10.0 - 35.0	23.3
SAMPLING DEGC	DEG C		
Sample ID			
Tested 08/24/23			
Tester CCOA			
Recorded 08/24/23			
DATE OF TESTING OF SAMPLE	MM/DD/YYYY		08/24/23
	MM/DD/YYYY		
Sample ID			
Tested 08/24/23			
Tester CCOA			
Recorded 08/24/23			
NORMALITY TO 5 DECIMAL PLACES	0.00100	0.00090 - 0.00120	0.00097
	EQUIV/LITER		
Sample ID			
Tested 08/24/23			
Tester CCOA			
Recorded 08/24/23			
Actual Density	0.79157	0.79066 - 0.79249	0.79179
	GM/ML		
Sample ID			
Tested 08/24/23			
Tester CCOA			
Recorded 08/24/23			
APPEARANCE AND COLOR	CLEAR LIQUID		PASS
	APPEARANCE		
Sample ID			
Tested 08/24/23			
Tester CCOA			
Recorded 08/24/23			



Aqua Solutions, Inc.
Ph: 281-479-2569
Fax: 281-479-2790
P.O. Box 70
Deer Park, TX 77536
Website: www.aquasolutions.org

Page: 2

Date: 08/24/23 at 1:32 PM

Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

INSTRUMENTS USED DURING PREPARATION		INSTRUMENT		M112,VF3-J21,M55,BID03828, PIDP40-5	
Sample ID					
Tested		08/24/23			
Tester		CCOA			
Recorded		08/24/23			
EXPIRATION DATE		MM/DD/YYYY		10/29/23	
		MM/DD/YYYY			
Sample ID					
Tested		08/24/23			
Tester		CCOA			
Recorded		08/24/23			
AMBIENT TEST TEMPERATURE		20.0		15.0 - 30.0	
DEG C		DEG C		23.5	
Sample ID					
Tested		08/24/23			
Tester		CCOA			
Recorded		08/24/23			
Lot #		3082137			
Made		08/24/23			

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

Quality Manager