

Page: 1

Date: 08/23/23 at 6:00 PM

Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results

Customer PO #:

Packaged Product	: 1793			
· ·		NE STANDARD 250 PPM AS C	L2	
Test		Target/UOM	Range	Result
PREPARED TO FORMU	JLATION ON	YES		YES
FILE		YES		
Sample ID				
Tested	08/23/23			
Tester	JENC			
Recorded	08/23/23			
APPEARANCE AND CO	LOR	CLEAR TO PALE YELLOW		PASS
		LIQUID APPEARANCE		
Sample ID				
Tested	08/23/23			
Tester	JENC			
Recorded	08/23/23			
TRACEABLE TO NIST SRM 136		YES		YES
		YES		
Sample ID				
Tested	08/23/23			
Tester	JENC			
Recorded	08/23/23			
INSTRUMENTS USED I	DURING	INSTRUMENT		M55,BID02259,PIDP-141
PREPARATION				
Sample ID				
Tested	08/23/23			
Tester	JENC			
Recorded	08/23/23			
Actual Density		0.99973	0.99923 - 1.00022	0.99984
		GM/ML		
Sample ID				
Tested	08/23/23			
Tester	JENC			
Recorded	08/23/23			
EXPIRATION DATE		MM/DD/YY		10/28/23
		MM/DD/YYYY		
Sample ID				
Tested	08/23/23			
Tester	JENC			
Recorded	08/23/23			
PPM CHLORINE (CL2)		250.0	246.0 - 254.0	250.5
		PPM		
Sample ID				
Tested	08/23/23			
Tester	JENC			
Recorded	08/23/23			
Lot # 3081	780			
Made 08/23	J / 4 J			



Page: 2

Date: 08/23/23 at 6:00 PM

Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO #:

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

Quality Manager