



Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

Packaged Product: 4090			
HYDROCHLORIC ACID 0.01 NORMAL			
Test	Target/UOM	Range	Result
PREPARED TO FORMULATION ON FILE	YES		YES
Sample ID			
Tested 08/22/23			
Tester JENC			
Recorded 08/22/23			
SAMPLE COLLECTION & HANDLING PER LAB040	YES		YES
Sample ID			
Tested 08/22/23			
Tester JENC			
Recorded 08/22/23			
DATE OF SAMPLING FOR TEST PURPOSES	MM/DD/YYYY		08/22/23
Sample ID			
Tested 08/22/23			
Tester JENC			
Recorded 08/22/23			
AMBIENT TEMPERATURE DURING SAMPLING DEGC	25.0 DEG C	10.0 - 35.0	23.4
Sample ID			
Tested 08/22/23			
Tester JENC			
Recorded 08/22/23			
DATE OF TESTING OF SAMPLE	MM/DD/YYYY MM/DD/YYYY		08/22/23
Sample ID			
Tested 08/22/23			
Tester JENC			
Recorded 08/22/23			
AMBIENT TEST TEMPERATURE DEGC	20.0 DEG C	15.0 - 30.0	23.4
Sample ID			
Tested 08/22/23			
Tester JENC			
Recorded 08/22/23			
Actual Normality	0.0100 EQ/L	0.0095 - 0.0105	0.0098
Sample ID			
Tested 08/22/23			
Tester JENC			
Recorded 08/22/23			
+ Measurement Uncertainty	0.0001 EQ/L	= 0.0001	0.0001
Sample ID			
Tested 08/22/23			
Tester JENC			
Recorded 08/22/23			



Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

QCSP5: STANDARDIZATION OF AQUEOUS ACIDS	YES METHOD		YES
Sample ID			
Tested	08/22/23		
Tester	JENC		
Recorded	08/22/23		
N.I.S.T. Traceable to SRM 723 THAM	YES YES		YES
Sample ID			
Tested	08/22/23		
Tester	JENC		
Recorded	08/22/23		
APPEARANCE AND COLOR	CLEAR LIQUID APPEARANCE		PASS
Sample ID			
Tested	08/22/23		
Tester	JENC		
Recorded	08/22/23		
Actual Density	0.99842 GM/ML	0.99832 - 0.99852	0.99839
Sample ID			
Tested	08/22/23		
Tester	JENC		
Recorded	08/22/23		
INSTRUMENTS USED DURING PREPARATION	INSTRUMENT		M112,T-243,M55,BID06419
Sample ID			
Tested	08/22/23		
Tester	JENC		
Recorded	08/22/23		
EXPIRATION DATE	MM/DD/YY MM/DD/YYYY		08/30/24
Sample ID			
Tested	08/22/23		
Tester	JENC		
Recorded	08/22/23		
SHELF LIFE OF PRODUCT	1 YEAR TIME		1 YEAR
Sample ID			
Tested	08/22/23		
Tester	JENC		
Recorded	08/22/23		

Lot # 3081018

Made 08/22/23



Aqua Solutions, Inc.
Ph: 281-479-2569
Fax: 281-479-2790
P.O. Box 70
Deer Park, TX 77536
Website: www.aquasolutions.org

Page: 3

Date: 08/22/23 at 4:08 PM

Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

Quality Manager