

Page: 1

Date: 08/24/23 at 4:12 PM

Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569

Certificate of Analysis / QC Results

Customer PO #:

Packaged Produc	t: CY157F			
_		STANDARD 100 PPM		
Test		Target/UOM	Range	Result
PREPARED TO FORM	ULATION ON	YES		YES
FILE		YES		
Sample ID)			
Tested	08/24/23			
Tester	MHUS			
Recorded	08/24/23			
Actual PPM		100.00	98.00 - 102.00	98.45
		PPM		
Sample ID)			
Tested	08/24/23			
Tester	MHUS			
Recorded				
QCSP47:STANDARDIZ		YES		YES
VARIOUS CYANIDE		QCSP47		,
Sample ID	1	Q001 11		
Tested	08/24/23			
Tester	MHUS			
Recorded				
APPEARANCE AND C		CLEAR LIQUID		PASS
APPEARANCE AND C	OLOR	APPEARANCE		FAGG
0		AFFEARANCE		
Sample ID				
Tested	08/24/23			
Tester	MHUS			
Recorded				
INSTRUMENTS USED	DURING	INSTRUMENT		M112,T44
PREPARATION				
Sample ID)			
Tested	08/24/23			
Tester	MHUS			
Recorded	08/24/23			
SHELF LIFE OF PRODUCT		6 WEEKS		6 WEEKS
		TIME		
Sample ID)			
Tested	08/24/23			
Tester	MHUS			
Recorded				
EXPIRATION DATE		MM/DD/YYYY		10/05/23
		MM/DD/YYYY		
Sample ID)			
Tested	08/24/23			
Tester	MHUS			
Recorded	08/24/23			
Made 08/2	4/23			

Page: 2

Date: 08/24/23 at 4:12 PM

Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO #:

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

Quality Manager