

**Aqua Solutions, Inc.**  
**6913 Highway 225**  
**Deer Park, TX, 77536**  
**281/479-2569**  
**Certificate of Analysis / QC Results**  
**Customer PO # :**

<b>Packaged Product:</b> 8667 SODIUM HYDROXIDE 1.2 NORMAL LOW CARBONATE			
Test	Target/UOM	Range	Result
<b>PREPARED TO FORMULATION ON</b>	YES		YES
<b>FILE</b>	YES		
Sample ID			
Tested 08/21/23			
Tester CCOA			
Recorded 08/21/23			
<b>SAMPLE COLLECTION &amp; HANDLING</b>	YES		YES
<b>PER LAB040</b>	YES		
Sample ID			
Tested 08/21/23			
Tester CCOA			
Recorded 08/21/23			
<b>DATE OF SAMPLING FOR TEST</b>	MM/DD/YYYY		08/21/23
<b>PURPOSES</b>	MM/DD/YYYY		
Sample ID			
Tested 08/21/23			
Tester CCOA			
Recorded 08/21/23			
<b>AMBIENT TEMPERATURE DURING</b>	25.0	10.0 - 35.0	28.3
<b>SAMPLING DEGC</b>	DEG C		
Sample ID			
Tested 08/21/23			
Tester CCOA			
Recorded 08/21/23			
<b>DATE OF TESTING OF SAMPLE</b>	MM/DD/YYYY		08/21/23
	MM/DD/YYYY		
Sample ID			
Tested 08/21/23			
Tester CCOA			
Recorded 08/21/23			
<b>AMBIENT TEST TEMPERATURE</b>	20.0	15.0 - 30.0	22.3
<b>DEG C</b>	DEG C		
Sample ID			
Tested 08/21/23			
Tester CCOA			
Recorded 08/21/23			
<b>Product nitrogen purged</b>	YES		YES
	YES		
Sample ID			
Tested 08/21/23			
Tester CCOA			
Recorded 08/21/23			
<b>QCSP7: STANDARDIZATION OF</b>	YES		YES
<b>AQUEOUS BASES</b>	METHOD		
Sample ID			
Tested 08/21/23			
Tester CCOA			
Recorded 08/21/23			

**Aqua Solutions, Inc.**  
**6913 Highway 225**  
**Deer Park, TX, 77536**  
**281/479-2569**  
**Certificate of Analysis / QC Results**  
**Customer PO # :**

<b>Actual Normality</b>	1.2000 EQ/L	1.1990 - 1.2010	1.2002
Sample ID			
Tested	08/21/23		
Tester	CCOA		
Recorded	08/21/23		
<b>N.I.S.T. Traceable to SRM 84 (KHP)</b>	YES YES		YES
Sample ID			
Tested	08/21/23		
Tester	CCOA		
Recorded	08/21/23		
<b>PPM INORGANIC CARBON</b>	25.0 PPM	0.0 - 25.0	7.6
Sample ID			
Tested	08/21/23		
Tester	CCOA		
Recorded	08/21/23		
<b>Actual Density</b>	1.04915 GM/ML	1.04815 - 1.05035	1.04906
Sample ID			
Tested	08/21/23		
Tester	CCOA		
Recorded	08/21/23		
<b>APPEARANCE AND COLOR</b>	CLEAR WATER WHITE LIQUID APPEARANCE		PASS
Sample ID			
Tested	08/21/23		
Tester	CCOA		
Recorded	08/21/23		
<b>INSTRUMENTS USED DURING PREPARATION</b>	INSTRUMENT		M55,T8667,M91,BID00488,TID 41086
Sample ID			
Tested	08/21/23		
Tester	CCOA		
Recorded	08/21/23		
<b>SHELF LIFE OF PRODUCT</b>	1 YEAR TIME		1 YEAR
Sample ID			
Tested	08/21/23		
Tester	CCOA		
Recorded	08/21/23		
<b>EXPIRATION DATE</b>	MM/DD/YYYY MM/DD/YYYY		08/30/24
Sample ID			
Tested	08/21/23		
Tester	CCOA		
Recorded	08/21/23		
<b>Lot #</b>	3081404		
<b>Made</b>	08/21/23		



Aqua Solutions, Inc.  
Ph: 281-479-2569  
Fax: 281-479-2790  
P.O. Box 70  
Deer Park, TX 77536  
Website: [www.aquasolutions.org](http://www.aquasolutions.org)

Page: 3

Date: 08/21/23 at 6:18 PM

**Aqua Solutions, Inc.**  
**6913 Highway 225**  
**Deer Park, TX, 77536**  
**281/479-2569**  
**Certificate of Analysis / QC Results**  
**Customer PO # :**

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

**Quality Manager**