



Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

Packaged Product: 8407			
SILVER NITRATE 0.025 NORMAL			
Test	Target/UOM	Range	Result
Actual Normality	0.0250 EQ/L	0.0245 - 0.0255	0.0249
Sample ID			
Tested	08/30/23		
Tester	MHUS		
Recorded	08/30/23		
QCSP53: SILVER NITRATE STD ON AUTO-TITRA	YES METHOD		YES
Sample ID			
Tested	08/30/23		
Tester	MHUS		
Recorded	08/30/23		
QCSP54: STANDARDIZATION OF LOW NORMAL SN	YES QCSP54		YES
Sample ID			
Tested	08/30/23		
Tester	MHUS		
Recorded	08/30/23		
QCSP78: CHLORIDE SQC METHOD TO VERIFY SN	YES QCSP78		YES
Sample ID			
Tested	08/30/23		
Tester	MHUS		
Recorded	08/30/23		
N.I.S.T. Traceable to SRM 919	Yes YES		YES
Sample ID			
Tested	08/30/23		
Tester	MHUS		
Recorded	08/30/23		
PREPARED TO FORMULATION ON FILE	YES YES		YES
Sample ID			
Tested	08/30/23		
Tester	MHUS		
Recorded	08/30/23		
SAMPLE COLLECTION & HANDLING PER LAB040	YES YES		YES
Sample ID			
Tested	08/30/23		
Tester	MHUS		
Recorded	08/30/23		
DATE OF SAMPLING FOR TEST PURPOSES	MM/DD/YYYY MM/DD/YYYY		08/30/2023
Sample ID			
Tested	08/30/23		
Tester	MHUS		
Recorded	08/30/23		



Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569

Certificate of Analysis / QC Results

Customer PO # :

AMBIENT TEMPERATURE DURING SAMPLING DEGC	25.0 DEG C	10.0 - 35.0	22.6
Sample ID			
Tested	08/30/23		
Tester	MHUS		
Recorded	08/30/23		
DATE OF TESTING OF SAMPLE	MM/DD/YYYY MM/DD/YYYY		08/30/2023
Sample ID			
Tested	08/30/23		
Tester	MHUS		
Recorded	08/30/23		
AMBIENT TEST TEMPERATURE DEGC	20.0 DEG C	15.0 - 30.0	22.6
Sample ID			
Tested	08/30/23		
Tester	MHUS		
Recorded	08/30/23		
APPEARANCE AND COLOR	CLEAR WATER WHITE LIQUID APPEARANCE		PASS
Sample ID			
Tested	08/30/23		
Tester	MHUS		
Recorded	08/30/23		
INSTRUMENTS USED DURING PREPARATION	INSTRUMENT		M55,VF5-E3,.M91
Sample ID			
Tested	08/30/23		
Tester	MHUS		
Recorded	08/30/23		
Actual Density	1.00181 GM/ML	1.00126 - 1.00236	1.00174
Sample ID			
Tested	08/30/23		
Tester	MHUS		
Recorded	08/30/23		
Lot #	3082868		
Made	08/30/23		

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.



Aqua Solutions, Inc.
Ph: 281-479-2569
Fax: 281-479-2790
P.O. Box 70
Deer Park, TX 77536
Website: www.aquasolutions.org

Page: 3

Date: 08/31/23 at 7:11 AM

Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

Quality Manager