

Page: 1

Date: 09/14/23 at 4:26 PM

## Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results

Customer PO #:

Packaged Product	t: 0369			
AMMONIA STANDARD 50.0 PPM AS NH3				
Test		Target/UOM	Range	Result
PREPARED TO FORMULATION ON		YES	<del>-</del>	YES
FILE		YES		
Sample ID	)			
Tested	09/14/23			
Tester	AROS			
Recorded	09/14/23			
CERTIFIED PPM NH3		50.0	48.0 - 52.0	48.1
		PPM		
Sample ID	)			
Tested	09/14/23			
Tester	AROS			
Recorded				
APPEARANCE AND COLOR		CLEAR LIQUID		PASS
		APPEARANCE		
Sample ID	)			
Tested	09/14/23			
Tester	AROS			
Recorded				
INSTRUMENTS USED		INSTRUMENT		VF100-K8,P143,GC,VF3-J1
PREPARATION	Domina	II TO THOME IT		VI 100 110, I 10, GG, VI 0 0 1
Sample ID	•			
Tested	09/14/23			
Tester	09/14/23 AROS			
Recorded		3 MONTHS		3 MONTHS
SHELF LIFE OF PRODUCT		TIME		3 MON 1 HS
0		TIME		
Sample ID				
Tested	09/14/23			
Tester	AROS			
Recorded	09/14/23	MM/DD 0000/		40/00/0000
EXPIRATION DATE		MM/DD/YYYY MM/DD/YYYY		12/30/2023
Sample ID	)			
Tested	09/14/23			
Tester	AROS			
Recorded	09/14/23			
Lot # 3091425				
Made 09/14/23				



Page: 2

Date: 09/14/23 at 4:26 PM

Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO #:

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

**Quality Manager**