



Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

Packaged Product: PPG036F			
COLOR REAGENT (CHLORIDES)			
Test	Target/UOM	Range	Result
PREPARED TO FORMULATION ON	YES		YES
FILE	YES		
Sample ID			
Tested 09/14/23			
Tester JRIO			
Recorded 09/14/23			
Final Product Filtered	YES		YES
	YES		
Sample ID			
Tested 09/14/23			
Tester JRIO			
Recorded 09/14/23			
Renex 30, 30% added to product	YES		YES
	YES		
Sample ID			
Tested 09/14/23			
Tester JRIO			
Recorded 09/14/23			
ACTUAL GRAMS OF MERCURIC THIOCYANATE	RECORD ACTUAL GMS M2275 GM		12.48
Sample ID			
Tested 09/14/23			
Tester JRIO			
Recorded 09/14/23			
TOTAL MLS OF NITRIC ACID ADDED	RECORD ACTUAL MLS NITRIC ACID MLS		1512
Sample ID			
Tested 09/14/23			
Tester JRIO			
Recorded 09/14/23			
RECORD ACTUAL GRAMS OF FERRIC NITRATE	RECORD ACT GMS FERRIC NITRATE GMS F2530		960
Sample ID			
Tested 09/14/23			
Tester JRIO			
Recorded 09/14/23			
FINAL VOLUME IN LITERS	RECORD FINAL VOLUME IN LITERS LITER		24
Sample ID			
Tested 09/14/23			
Tester JRIO			
Recorded 09/14/23			



Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

APPEARANCE AND COLOR	BROWN LIQUID APPEARANCE	PASS
Sample ID		
Tested	09/14/23	
Tester	JRIO	
Recorded	09/14/23	
INSTRUMENTS USED DURING PREPARATION	INSTRUMENT	M115,VF4-N19,GC,M103
Sample ID		
Tested	09/14/23	
Tester	AROS	
Recorded	09/14/23	
VERIFICATION	PASS PASS	PASS
Sample ID		
Tested	09/14/23	
Tester	JRIO	
Recorded	09/14/23	
RECORD ACTUAL MLS OF METHANOL USED	RECORD ACTUAL MLS OF METHANOL MLS	4800
Sample ID		
Tested	09/14/23	
Tester	JRIO	
Recorded	09/20/23	
Lot #	3091342	
Made	09/14/23	

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

Quality Manager