

Page: 1

Date: 09/05/23 at 1:51 PM

Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569

Certificate of Analysis / QC Results

Customer PO #:

Packaged Product	t: 0371			
		Nitrogen Standard 10 F	PPM as NH3	
Test		Target/UOM	Range	Result
PREPARED TO FORMULATION ON		YES		YES
FILE		YES		
Sample ID)			
Tested	09/05/23			
Tester	AROS			
Recorded	09/05/23			
CERTIFIED PPM NH3		10.0	9.0 - 11.0	10.0
		PPM		
Sample ID)			
Tested	09/05/23			
Tester	AROS			
Recorded				
APPEARANCE AND COLOR		CLEAR LIQUID		PASS
		APPEARANCE		
Sample ID)			
Tested	09/05/23			
Tester	AROS			
Recorded	09/05/23			
INSTRUMENTS USED		INSTRUMENT		P-143,VF5-E3
PREPARATION				-, -
Sample ID	1			
Tested	09/05/23			
Tester	AROS			
Recorded	09/05/23			
EXPIRATION DATE	03/03/20	MM/DD/YY		12/30/23
EXPINATION DATE		MM/DD/YYYY		12/00/20
Sample ID	•	IVIIVI/DD/TTTT		
Tested	09/05/23			
Tested	09/05/23 AROS			
r ester Recorded				
SHELF LIFE OF PROD		SHELF LIFE		3 MONTHS
SHELF LIFE OF PROD	UUI	TIME		S INIOIN I US
Sample ID)			
Tested	09/05/23			
Tester	AROS			
Recorded	09/05/23			
Lot # 3090547				
Made 09/05/23				
IVIAGE 537 6	-,			



Page: 2

Date: 09/05/23 at 1:51 PM

Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO #:

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

Quality Manager