



**Aqua Solutions, Inc.**  
**6913 Highway 225**  
**Deer Park, TX, 77536**  
**281/479-2569**  
**Certificate of Analysis / QC Results**  
**Customer PO # :**

<b>Packaged Product:</b> LY386F			
SULFUR STANDARD 10.0 PPM W/W IN ISOOCTANE			
Test	Target/UOM	Range	Result
<b>PREPARED TO FORMULATION ON</b>	YES		YES
<b>FILE</b>	YES		
Sample ID			
Tested 09/08/23			
Tester MMAT			
Recorded 09/08/23			
<b>Calculated ppm Sulfur</b>	10.00	9.80 - 10.20	10.00
	PPM		
Sample ID			
Tested 09/08/23			
Tester MMAT			
Recorded 09/08/23			
<b>APPEARANCE AND COLOR</b>	CLEAR LIQUID APPEARANCE		PASS
Sample ID			
Tested 09/08/23			
Tester MMAT			
Recorded 09/08/23			
<b>INSTRUMENTS USED DURING PREPARATION</b>	INSTRUMENT		P125,VF1-Z40
Sample ID			
Tested 09/08/23			
Tester MMAT			
Recorded 09/08/23			
<b>Lot #</b>	3090863		
<b>Made</b>	09/08/23		

**Quality Manager**