

Page: 1

Date: 09/14/23 at 2:09 PM

Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569

Certificate of Analysis / QC Results
Customer PO #:

Packaged Product:				
	CONDUC	TIVITY STANDARD 2,000 uMM		
Test		Target/UOM	Range	Result
PREPARED TO FORMU	LATION ON	YES		YES
FILE		YES		
Sample ID				
Tested	09/14/23			
Tester	RSHE			
Recorded	09/14/23			
CONDUCTIVITY @25 C.		2000	1999 - 2001	2000
		US/CM		
Sample ID				
Tested	09/14/23			
Tester	RSHE			
Recorded	09/14/23			
QCSP30:CONDUCTIVIT	Υ	YES		YES
STD&MEASUREMENT		QCSP30		
Sample ID				
Tested	09/14/23			
Tester	RSHE			
Recorded	09/14/23			
TRACEABLE TO N.I.S.T	. SRM 3191	YES		YES
		YES		
Sample ID				
Tested	09/14/23			
Tester	RSHE			
Recorded	09/14/23			
APPEARANCE AND COLOR		CLEAR WATER WHITE		PASS
		LIQUID		
		APPEARANCE		
Sample ID				
Tested	09/14/23			
Tester	RSHE			
Recorded	09/14/23			
INSTRUMENTS USED D	URING	INSTRUMENT		T-229
PREPARATION				
Sample ID				
Tested	09/14/23			
Tester	RSHE			
Recorded	09/14/23			
EXPIRATION DATE		MM/DD/YYYY		09/30/2024
		MM/DD/YYYY		
Sample ID				
Tested	09/14/23			
Tester	RSHE			
Recorded	09/14/23			
Lot # 30914	11			
Made 09/14				
IVIAUE 03/14	, 10			



Page: 2

Date: 09/14/23 at 2:09 PM

Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO #:

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

Quality Manager