



Aqua Solutions, Inc.
Ph: 281-479-2569
Fax: 281-479-2790
P.O. Box 70
Deer Park, TX 77536
Website: www.aquasolutions.org

Page: 1

Date: 09/08/23 at 9:10 AM

Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

Packaged Product: SPX470F Water Standard 10%w/w (8 x 10 ml)			
Test	Target/UOM	Range	Result
PREPARED TO FORMULATION ON	YES		YES
FILE	YES		
Sample ID			
Tested 09/07/23			
Tester RIRO			
Recorded 09/07/23			
W/W% WATER	10.00 W/W%	9.80 - 10.20	9.96
Sample ID			
Tested 09/07/23			
Tester RIRO			
Recorded 09/07/23			
APPEARANCE AND COLOR	CLEAR LIQUID APPEARANCE		PASS
Sample ID			
Tested 09/07/23			
Tester RIRO			
Recorded 09/07/23			
INSTRUMENTS USED DURING PREPARATION	INSTRUMENT		M112,EF1-PLD
Sample ID			
Tested 09/07/23			
Tester RIRO			
Recorded 09/07/23			
Lot #	3081627		
Made	09/08/23		

Quality Manager