09/05/23

Made

Page: 1

Date: 09/05/23 at 4:55 PM

## Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results

Customer PO #:

Packag	ed Product	DEL180F			
		COMBINE	O ANION STD.		
Test			Target/UOM	Range	Result
PREPAR	ED TO FORMU	ILATION ON	YES	-	YES
FILE			YES		
	Sample ID				
	Tested	09/05/23			
	Tester	AROS			
	Recorded	09/05/23			
APPEARANCE AND COLO		LOR	CLEAR LIQUID		PASS
			APPEARANCE		
	Sample ID				
	Tested	09/05/23			
	Tester	AROS			
	Recorded	09/05/23			
INSTRUMENTS USED DURING			INSTRUMENT		P-385,P-93,P-37,VF2-B15
PREPAR	ATION				
	Sample ID				
	Tested	09/05/23			
	Tester	AROS			
	Recorded	09/05/23			
EXPIRATION DATE			MM/DD/YYYY		03/30/2024
			MM/DD/YYYY		
	Sample ID				
	Tested	09/05/23			
	Tester	AROS			
	Recorded	09/05/23			
L	ot# 30905	519			

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

**Quality Manager** 

hichard Stot