



Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

| Packaged Product: FIS107F | | | |
|---|-------------------|---------|------------|
| Conductivity Standard 74.0 umhos @25°C | | | |
| Test | Target/UOM | Range | Result |
| PREPARED TO FORMULATION ON FILE | YES | | YES |
| Sample ID | YES | | |
| Tested 09/12/23 | | | |
| Tester RSHE | | | |
| Recorded 09/12/23 | | | |
| CONDUCTIVITY @25 C. | 74 | 73 - 75 | 74 |
| | US/CM | | |
| Sample ID | | | |
| Tested 09/12/23 | | | |
| Tester RSHE | | | |
| Recorded 09/12/23 | | | |
| QCSP30: CONDUCTIVITY STD&MEASUREMENT | YES | | YES |
| | QCSP30 | | |
| Sample ID | | | |
| Tested 09/12/23 | | | |
| Tester RSHE | | | |
| Recorded 09/12/23 | | | |
| TRACEABLE TO N.I.S.T. SRM 3191 | YES | | YES |
| | YES | | |
| Sample ID | | | |
| Tested 09/12/23 | | | |
| Tester RSHE | | | |
| Recorded 09/12/23 | | | |
| APPEARANCE AND COLOR | CLEAR WATER WHITE | | PASS |
| | APPEARANCE | | |
| Sample ID | | | |
| Tested 09/12/23 | | | |
| Tester RSHE | | | |
| Recorded 09/12/23 | | | |
| INSTRUMENTS USED DURING PREPARATION | INSTRUMENT | | M92,T234 |
| Sample ID | | | |
| Tested 09/12/23 | | | |
| Tester RSHE | | | |
| Recorded 09/12/23 | | | |
| EXPIRATION DATE | MM/DD/YYYY | | 09/30/2024 |
| | MM/DD/YYYY | | |
| Sample ID | | | |
| Tested 09/12/23 | | | |
| Tester RSHE | | | |
| Recorded 09/12/23 | | | |
| Lot # | 3091250 | | |
| Made | 09/12/23 | | |



Aqua Solutions, Inc.
Ph: 281-479-2569
Fax: 281-479-2790
P.O. Box 70
Deer Park, TX 77536
Website: www.aquasolutions.org

Page: 2

Date: 09/12/23 at 2:50 PM

Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

Quality Manager