

Page: 1

Date: 09/12/23 at 4:02 PM

Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569

Certificate of Analysis / QC Results
Customer PO #:

D I	I Door doors	SPX190F			
Package	d Product:				
		HYDROGEN		W SOLUTION CERTIFIED	
Test			Target/UOM	Range	Result
PREPARED TO FORMULATION ON		YES		YES	
FILE			YES		
	Sample ID				
	Tested	09/12/23			
	Tester	MHUS			
	Recorded	09/12/23			
PPM PEROXIDE			500.0	480.0 - 520.0	509.5
			PPM		
	Sample ID				
	Tested	09/12/23			
	Tester	MHUS			
	Recorded	09/12/23			
Actual Density			0.99840	0.99831 - 0.99849	0.99846
			GM/ML		
	Sample ID				
	Tested	09/12/23			
	Tester	MHUS			
	Recorded	09/12/23			
APPEARANCE AND COLOR		LOR	CLEAR LIQUID		PASS
			APPEARANCE		
	Sample ID				
	Tested	09/12/23			
	Tester	MHUS			
	Recorded	09/12/23			
INSTRUMENTS USED DURING		INSTRUMENT		M112,VF2-B3,M55,BID04099,P	
PREPARATION				IDM91	
	Sample ID				
	Tested	09/12/23			
	Tester	MHUS			
	Recorded	09/12/23			
EXPIRATION DATE			MM/DD/YYYY		10/24/2023
			MM/DD/YYYY		
	Sample ID				
	Tested	09/12/23			
	Tester	MHUS			
	Recorded	09/12/23			
I ot	# 30908	11			
Made 09/12/23					
iviac	16 02/12	, 20			



Page: 2

Date: 09/12/23 at 4:02 PM

Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO #:

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

Quality Manager