

Page: 1

Date: 09/08/23 at 11:24 AM

## Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results

Customer PO #:

Packaged Product	TEN222	F		
		E INDEX STANDARD 18 - 20	MG/100G (1 LITER FORMULA	ATION)
Test		Target/UOM	Range	Result
PREPARED TO FORMULATION ON		YES		YES
FILE		YES		
Sample ID				
Tested	09/08/23			
Tester	MMAT			
Recorded	09/08/23			
BROMINE INDEX VALUE		19.00	15.00 - 24.00	18.30
		BROMINE IN		
Sample ID				
Tested	09/08/23			
Tester	MMAT			
Recorded	09/08/23			
APPEARANCE AND COLOR		CLEAR WATER WHITE		PASS
		APPEARANCE		
Sample ID				
Tested	09/08/23			
Tester	MMAT			
Recorded	09/08/23			
INSTRUMENTS USED		LAB EQUIPMENT USED		M100,M115
PREPARATION		INSTRUMENT		,
Sample ID				
Tested	09/08/23			
Tester	MMAT			
Recorded	09/08/23			
EXPIRATION DATE	00,00,20	MM/DD/YY		12/30/23
		MM/DD/YYYY		
Sample ID				
Tested	09/08/23			
Tester	MMAT			
Recorded	09/08/23			
SHELF LIFE OF PROD		3 MONTHS		3 MONTHS
		TIME		5 <b>5</b> 111115
Sample ID				
Tested	09/08/23			
Tester	MMAT			
Recorded	09/08/23			
<b>Lot #</b> 3090				
<b>Made</b> 09/0				
indue 557 G				

Page: 2

Date: 09/08/23 at 11:24 AM

Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO #:

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

**Quality Manager**