

Page: 1

Date: 09/06/23 at 5:48 PM

Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results

Customer PO #:

Packaged Product: 1764 CHLORIDE STANDARD 1000 PPM as Cl Certified				
Test		Target/UOM	Range	Result
PREPARED TO FORMU	LATION ON	YES	-	YES
FILE		YES		
Sample ID				
Tested	09/06/23			
Tester	RIRO			
Recorded	09/06/23			
CHLORIDE PPM		1000.0 PPM	980.0 - 1020.0	1002.2
Sample ID				
Tested	09/06/23			
Tester	RIRO			
Recorded	09/06/23			
N.I.S.T. Traceable to SR		YES		YES
	W 515 14051	YES		120
Sample ID				
Tested	09/06/23			
Tester	RIRO			
Recorded	09/06/23			
APPEARANCE AND CO	LOR	CLEAR LIQUID APPEARANCE		PASS
Sample ID				
Tested	09/06/23			
Tester	RIRO			
Recorded	09/06/23			
INSTRUMENTS USED D	URING	INSTRUMENT		M91,VF3-J16
PREPARATION				
Sample ID				
Tested	09/06/23			
Tester	RIRO			
Recorded	09/06/23			
EXPIRATION DATE		MM/DD/YY		09/30/24
		MM/DD/YYYY		
Sample ID	00/06/22			
Tested	09/06/23			
Tester	RIRO			
Recorded	09/06/23	1 YEAR		4 \/EAD
SHELF LIFE OF PRODU	IC I	TIME		1 YEAR
Sample ID				
Tested	09/06/23			
Tester	RIRO			
Recorded	09/06/23			



Page: 2

Date: 09/06/23 at 5:48 PM

Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO #:

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

Quality Manager