



Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

Packaged Product: FIS108F			
Conductivity Standard 149.6 umhos @25°C			
Test	Target/UOM	Range	Result
PREPARED TO FORMULATION ON	YES		YES
FILE	YES		
Sample ID			
Tested 09/13/23			
Tester RSHE			
Recorded 09/13/23			
CONDUCTIVITY 1 DEC	149.6	149.5 - 149.7	149.6
Sample ID			
Tested 09/13/23			
Tester RSHE			
Recorded 09/13/23			
QCSP30: CONDUCTIVITY	YES		YES
STD&MEASUREMENT	QCSP30		
Sample ID			
Tested 09/13/23			
Tester RSHE			
Recorded 09/13/23			
TRACEABLE TO N.I.S.T. SRM 3191	YES		YES
	YES		
Sample ID			
Tested 09/13/23			
Tester RSHE			
Recorded 09/13/23			
APPEARANCE AND COLOR	CLEAR WATER WHITE		PASS
	APPEARANCE		
Sample ID			
Tested 09/13/23			
Tester RSHE			
Recorded 09/13/23			
INSTRUMENTS USED DURING	INSTRUMENT		T229,M92
PREPARATION			
Sample ID			
Tested 09/13/23			
Tester RSHE			
Recorded 09/13/23			
EXPIRATION DATE	MM/DD/YYYY		09/30/2024
	MM/DD/YYYY		
Sample ID			
Tested 09/13/23			
Tester RSHE			
Recorded 09/13/23			
Lot #	3091251		
Made	09/13/23		



Aqua Solutions, Inc.
Ph: 281-479-2569
Fax: 281-479-2790
P.O. Box 70
Deer Park, TX 77536
Website: www.aquasolutions.org

Page: 2

Date: 09/13/23 at 4:10 PM

Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

Quality Manager