

**Aqua Solutions, Inc.**  
**6913 Highway 225**  
**Deer Park, TX, 77536**  
**281/479-2569**  
**Certificate of Analysis / QC Results**  
**Customer PO # :**

<b>Packaged Product:</b> 0367			
AMMONIA STANDARD 1 PPM AS NH3			
<b>Test</b>	<b>Target/UOM</b>	<b>Range</b>	<b>Result</b>
<b>PREPARED TO FORMULATION ON</b>	YES		YES
<b>FILE</b>	YES		
Sample ID			
Tested 09/14/23			
Tester AROS			
Recorded 09/14/23			
<b>CERTIFIED PPM NH3</b>	1.0 PPM	0.8 - 1.2	0.9
Sample ID			
Tested 09/14/23			
Tester AROS			
Recorded 09/14/23			
<b>APPEARANCE AND COLOR</b>	CLEAR LIQUID APPEARANCE		PASS
Sample ID			
Tested 09/14/23			
Tester AROS			
Recorded 09/14/23			
<b>INSTRUMENTS USED DURING PREPARATION</b>	INSTRUMENT		P352,GC,VF3-J19,M107
Sample ID			
Tested 09/14/23			
Tester AROS			
Recorded 09/14/23			
<b>EXPIRATION DATE</b>	MM/DD/YYYY MM/DD/YYYY		11/19/2023
Sample ID			
Tested 09/14/23			
Tester AROS			
Recorded 09/14/23			
<b>Lot #</b>	3091424		
<b>Made</b>	09/14/23		

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

