

Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results

Customer PO # :

Packaged Product: ERL125F			
-	NITRATE 0.1 NORMAL IN	ISOPROPANOL	
Test	Target/UOM	Range	Result
PREPARED TO FORMULATION ON	YES		YES
FILE	YES		
Sample ID			
Tested 09/06/23			
Tester MHUS			
Recorded 09/06/23			
SAMPLE COLLECTION & HANDLING	YES		YES
PER LAB040	YES		
Sample ID			
Tested 09/06/23			
Tester MHUS			
Recorded 09/06/23			
DATE OF SAMPLING FOR TEST	MM/DD/YYYY		09/06/2023
PURPOSES	MM/DD/YYYY		
Sample ID			
Tested 09/06/23			
Tester MHUS			
Recorded 09/06/23			
AMBIENT TEMPERATURE DURING	25.0	10.0 - 35.0	22.0
SAMPLING DEGC	DEG C		
Sample ID			
Tested 09/06/23			
Tester MHUS			
Recorded 09/06/23			
DATE OF TESTING OF SAMPLE	MM/DD/YYYY		09/06/2023
	MM/DD/YYYY		
Sample ID			
Tested 09/06/23			
Tester MHUS			
Recorded 09/06/23			
AMBIENT TEST TEMPERATURE	20.0	15.0 - 30.0	22.0
DEG C	DEG C		
Sample ID			
Tested 09/06/23			
Tester MHUS			
Recorded 09/06/23			
Actual Normality	0.1000	0.0995 - 0.1005	0.0983
· · · · · · · · · · · · · · · · · · ·	EQ/L		
Sample ID			
Tested 09/06/23			
Tester MHUS			
Recorded 09/06/23			
QCSP53: SILVER NITRATE STD ON	YES		YES
AUTO-TITRA	METHOD		
Sample ID			
Tested 09/06/23			
Tester MHUS			
Recorded 09/06/23			



Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results

Customer PO # :

N.I.S.T. Traceable to SRM 9	919 NaCl	YES		YES
		YES		
Sample ID				
	9/06/23			
Tester M	HUS			
	9/06/23			
PREPARED TO ASTM D322	27	YES		YES
		YES		
Sample ID				
Tested 09	0/06/23			
Tester M	HUS			
	9/06/23			
Actual Density		0.82606	0.82377 - 0.82832	0.82725
		GM/ML		
Sample ID				
Tested 09	9/06/23			
Tester M	HUS			
Recorded 09	0/06/23			
APPEARANCE AND COLOR	र	CLEAR LIQUID		PASS
		APPEARANCE		
Sample ID				
	0/06/23			
	HUS			
Recorded 09	0/06/23			
INSTRUMENTS USED DUR	ING	INSTRUMENT		M55,P41,BID41123
PREPARATION				
Sample ID				
	0/06/23			
Tester M	HUS			
Recorded 09	0/06/23			
LOT NUMBER OF ORIGINA	L	LOT #		2111807
MATERIAL				
Sample ID				
	9/06/23			
Tester M	HUS			
Recorded 09	0/06/23			
EXPIRATION DATE		MM/DD/YY		09/19/23
		MM/DD/YYYY		
Sample ID				
	9/06/23			
	HUS			
	0/06/23			
SHELF LIFE OF PRODUCT		1 WEEK		1 WEEK
		TIME		
Sample ID				
-	9/06/23			
	HUS			
	0/06/23			
Lot # 3090105				
Made 09/06/2	5			



Page: 3 Date: 09/06/23 at 4:51 PM

Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results Customer PO # :

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

ichard Stop

Quality Manager