



**Aqua Solutions, Inc.**  
**6913 Highway 225**  
**Deer Park, TX, 77536**  
**281/479-2569**  
**Certificate of Analysis / QC Results**  
**Customer PO # :**

| <b>Packaged Product:</b> ERL114AF                  |                 |                 |            |
|--|-----------------|-----------------|------------|
| POTASSIUM HYDROXIDE 0.1 NORMAL IN IPA              |                 |                 |            |
| Test   | Target/UOM      | Range           | Result     |
| <b>PREPARED TO FORMULATION ON FILE</b>             | YES             |                 | YES        |
| Sample ID  |                 |                 |            |
| Tested 09/07/23                                    |                 |                 |            |
| Tester JENC  |                 |                 |            |
| Recorded 09/07/23                                  |                 |                 |            |
| <b>SAMPLE COLLECTION &amp; HANDLING PER LAB040</b> | YES             |                 | YES        |
| Sample ID  |                 |                 |            |
| Tested 09/07/23                                    |                 |                 |            |
| Tester JENC  |                 |                 |            |
| Recorded 09/07/23                                  |                 |                 |            |
| <b>DATE OF SAMPLING FOR TEST PURPOSES</b>          | MM/DD/YYYY      |                 | 09/07/2023 |
| Sample ID  |                 |                 |            |
| Tested 09/07/23                                    |                 |                 |            |
| Tester JENC  |                 |                 |            |
| Recorded 09/07/23                                  |                 |                 |            |
| <b>AMBIENT TEMPERATURE DURING SAMPLING DEGC</b>    | 25.0            | 10.0 - 35.0     | 22.8       |
| Sample ID  |                 |                 |            |
| Tested 09/07/23                                    |                 |                 |            |
| Tester JENC  |                 |                 |            |
| Recorded 09/07/23                                  |                 |                 |            |
| <b>DATE OF TESTING OF SAMPLE</b>                   | MM/DD/YYYY      |                 | 09/07/2023 |
| Sample ID  |                 |                 |            |
| Tested 09/07/23                                    |                 |                 |            |
| Tester JENC  |                 |                 |            |
| Recorded 09/07/23                                  |                 |                 |            |
| <b>AMBIENT TEST TEMPERATURE DEGC</b>               | 20.0            | 15.0 - 30.0     | 22.8       |
| Sample ID  |                 |                 |            |
| Tested 09/07/23                                    |                 |                 |            |
| Tester JENC  |                 |                 |            |
| Recorded 09/07/23                                  |                 |                 |            |
| <b>LOT NUMBER OF ORIGINAL MATERIAL</b>             | BULK LOT NUMBER |                 | 3081405    |
| Sample ID  | LOT #           |                 |            |
| Tested 09/07/23                                    |                 |                 |            |
| Tester JENC  |                 |                 |            |
| Recorded 09/07/23                                  |                 |                 |            |
| <b>Actual Normality</b>                            | 0.1000          | 0.0995 - 0.1005 | 0.1003     |
| Sample ID  | EQ/L            |                 |            |
| Tested 09/07/23                                    |                 |                 |            |
| Tester JENC  |                 |                 |            |
| Recorded 09/07/23                                  |                 |                 |            |



**Aqua Solutions, Inc.**  
**6913 Highway 225**  
**Deer Park, TX, 77536**  
**281/479-2569**  
**Certificate of Analysis / QC Results**  
**Customer PO # :**

|  |   |                   |                           |
|--|---|-------------------|---------------------------|
| <b>Actual Density</b>                      | 0.79116<br>GM/ML                            | 0.78994 - 0.79237 | 0.79131                   |
| Sample ID                                  |   |                   |                           |
| Tested                                     | 09/07/23                                    |                   |                           |
| Tester                                     | JENC  |                   |                           |
| Recorded                                   | 09/07/23                                    |                   |                           |
| <b>N.I.S.T. Traceable to SRM 84 (KHP)</b>  | YES<br>YES                                  |                   | YES                       |
| Sample ID                                  |   |                   |                           |
| Tested                                     | 09/07/23                                    |                   |                           |
| Tester                                     | JENC  |                   |                           |
| Recorded                                   | 09/07/23                                    |                   |                           |
| <b>APPEARANCE AND COLOR</b>                | CLEAR-SLIGHTLY HAZY<br>LIQUID<br>APPEARANCE |                   | PASS                      |
| Sample ID                                  |   |                   |                           |
| Tested                                     | 09/07/23                                    |                   |                           |
| Tester                                     | JENC  |                   |                           |
| Recorded                                   | 09/07/23                                    |                   |                           |
| <b>INSTRUMENTS USED DURING PREPARATION</b> | LAB EQUIPMENT<br>INSTRUMENT                 |                   | M55,M91,BID02259,TID41100 |
| Sample ID                                  |   |                   |                           |
| Tested                                     | 09/07/23                                    |                   |                           |
| Tester                                     | JENC  |                   |                           |
| Recorded                                   | 09/07/23                                    |                   |                           |
| <b>EXPIRATION DATE</b>                     | MM/DD/YYYY<br>MM/DD/YYYY                    |                   | 09/19/2023                |
| Sample ID                                  |   |                   |                           |
| Tested                                     | 09/07/23                                    |                   |                           |
| Tester                                     | JENC  |                   |                           |
| Recorded                                   | 09/07/23                                    |                   |                           |
| <b>SHELF LIFE OF PRODUCT</b>               | 1 WEEK<br>TIME                              |                   | 1 WEEK                    |
| Sample ID                                  |   |                   |                           |
| Tested                                     | 09/07/23                                    |                   |                           |
| Tester                                     | JENC  |                   |                           |
| Recorded                                   | 09/07/23                                    |                   |                           |
| <b>Lot #</b>                               | 3090104                                     |                   |                           |
| <b>Made</b>                                | 09/07/23                                    |                   |                           |

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.



Aqua Solutions, Inc.  
Ph: 281-479-2569  
Fax: 281-479-2790  
P.O. Box 70  
Deer Park, TX 77536  
Website: [www.aquasolutions.org](http://www.aquasolutions.org)

Page: 3

Date: 09/07/23 at 10:16 AM

**Aqua Solutions, Inc.**  
**6913 Highway 225**  
**Deer Park, TX, 77536**  
**281/479-2569**  
**Certificate of Analysis / QC Results**  
**Customer PO # :**

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

**Quality Manager**