**Made** 09/29/23

Page: 1

Date: 09/29/23 at 4:20 PM

## Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569

## Certificate of Analysis / QC Results

Customer PO #:

Packag	ed Product: 6487			
	NITRI	TE STANDARD 300 PPM (NO2)	_	
Test		Target/UOM	Range	Result
PREPARED TO FORMULATION ON		YES		YES
FILE		YES		
	Sample ID			
	Tested 09/29/23			
	Tester AROS			
	Recorded 09/29/23			
PPM NITRITE AS NO2		300.0	296.0 - 304.0	299.8
		PPM		
	Sample ID			
	Tested 09/29/23			
	Tester AROS			
	Recorded 09/29/23			
APPEARANCE AND COLOR		CLEAR WATER WHITE		PASS
		LIQUID		
		APPEARANCE		
	Sample ID	70 1 270 0 000		
	Tested 09/29/23			
	Tester AROS			
	Recorded 09/29/23			
NETDIIN	MENTS USED DURING	INSTRUMENT		M108,VF500-X42
PREPAR		INSTITUTION		1V1100, V1 300-X42
TREPAR				
	Sample ID			
	Tested 09/29/23			
	Tester AROS			
	Recorded 09/29/23	111/22000/		0.4 (0.0 (0.0 = -
EXPIRATION DATE		MM/DD/YYYY		01/30/2024
		MM/DD/YYYY		
	Sample ID			
	Tested 09/29/23			
	Tester AROS			
	Recorded 09/29/23			

**Quality Manager**