

Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results

Customer PO # :

Deelvened	Due due t	UCT126F			
Packaged I	-roduct:				
		CELITE	STD FOR TSS 40 MG/L SOLUTI	ON	
Test			Target/UOM	Range	Result
PREPARED T	O FORMUL	ATION ON	YES		YES
FILE			YES		
S	Sample ID				
Т	ested	09/18/23			
Т	ester	AROS			
F	Recorded	09/18/23			
ACTUAL PPM			40.00	39.50 - 40.50	40.03
			PPM		
S	Sample ID				
Т	ested	09/18/23			
Т	ester	AROS			
F	Recorded	09/18/23			
APPEARANCE AND COLOR		OR	SLIGHTLY TURBID LIQUID		PASS
			APPEARANCE		
S	Sample ID				
Т	ested	09/18/23			
Т	ester	AROS			
F	Recorded	09/18/23			
INSTRUMENTS USED DURING		JRING	INSTRUMENT		M115,VF4-N12
PREPARATIO	Ν				
S	Sample ID				
Т	ested	09/18/23			
Т	ester	AROS			
F	Recorded	09/18/23			
EXPIRATION	DATE		MM/DD/YY		03/30/24
			MM/DD/YYYY		
S	Sample ID				
Т	ested	09/18/23			
Т	ester	AROS			
F	Recorded	09/18/23			
Lot #	309180	09			
Made	09/18/	/23			
inade					

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

huland SPA