

Page: 1

Date: 09/25/23 at 4:48 PM

Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results

Customer PO #:

Packaged Product: HON136F				
TISAB Buffer Special			_	
Test		Target/UOM	Range	Result
PREPARED TO FORMULATION ON		YES		YES
FILE		YES		
	Sample ID			
	Tested 09/25/23			
	Tester AFRE			
	Recorded 09/25/23			
APPEARA	ANCE AND COLOR	CLEAR WATER WHITE		PASS
		APPEARANCE		
	Sample ID			
	Tested 09/25/23			
	Tester AFRE			
	Recorded 09/25/23			
INSTRUM	IENTS USED DURING	INSTRUMENT		VF6-19,GC,M113,M112,T243
PREPARA	ATION			
	Sample ID			
	Tested 09/25/23			
	Tester AFRE			
	Recorded 09/25/23			
Actual pH	l of product	5.300	5.000 - 5.500	5.335
riotaai pi	. o. p. oddol	PH	0.000	0.000
	Sample ID			
	Tester AFRE			
=VDID 4 =	Recorded 09/25/23	NANA/DD DOO		00/00/04
EXPIRATI	ION DATE	MM/DD/YY		09/30/24
		MM/DD/YYYY		
	Sample ID			
	Tested 09/25/23			
	Tester AFRE			
	Recorded 09/25/23			
SHELF LIFE OF PRODUCT		1 YEAR		1 YEAR
		TIME		
	Sample ID			
	Tested 09/25/23			
	Tester AFRE			
	Recorded 09/25/23			
FLOURID	E LEVEL	0.2	0.0 - 0.5	0.1
		PPM		
	Sample ID			
	Tested 09/25/23			
	Tester AFRE			
	Recorded 09/25/23			
L	ot # 3092058 ade 09/25/23			



Page: 2

Date: 09/25/23 at 4:48 PM

Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO #:

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

Quality Manager