

Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results

Customer PO # :

Packaged	Product:	0370			
		AMMONIA	STOCK STANDARD 500 pp	m AS NH3	
Test			Target/UOM	Range	Result
PREPARED	TO FORMUL	ATION ON	YES		YES
FILE			YES		
	Sample ID				
		09/19/23			
	Tester	AROS			
	Recorded	09/19/23			
CERTIFIED	PPM NH3		500.0	490.0 - 510.0	505.7
			PPM		
	Sample ID				
	Tested	09/19/23			
	Tester	AROS			
	Recorded	09/19/23			
APPEARANCE AND COLOR		.OR	CLEAR LIQUID		PASS
			APPEARANCE		
	Sample ID				
	Tested	09/19/23			
	Tester	AROS			
	Recorded	09/19/23			
INSTRUMENTS USED DURING		JRING	INSTRUMENT		M108,GC,VF2-E3
PREPARATI	ON				
	Sample ID				
	Tested	09/19/23			
	Tester	AROS			
	Recorded	09/19/23			
EXPIRATION DATE			MM/DD/YYYY		12/30/2023
			MM/DD/YYYY		
	Sample ID				
		09/19/23			
		AROS			
		09/19/23			
Lot	# 309192	24			
Made					
maa					

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

huland SPA