

## Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results

## Customer PO # :

Packaged Produc	CY157F			
	CYANIDE	STANDARD 100 PPM		
Test		Target/UOM	Range	Result
PREPARED TO FORM	ULATION ON	YES		YES
FILE		YES		
Sample ID				
Tested	09/19/23			
Tester	MHUS			
Recorded	09/19/23			
Actual PPM		100.00	98.00 - 102.00	100.34
		PPM		
Sample ID	)			
Tested	09/19/23			
Tester	MHUS			
Recorded	09/19/23			
QCSP47:STANDARDIZ	ATION	YES		YES
VARIOUS CYANIDE		QCSP47		
Sample ID	)			
Tested	09/19/23			
Tester	MHUS			
Recorded	09/19/23			
APPEARANCE AND CO	OLOR	CLEAR LIQUID		PASS
		APPEARANCE		
Sample ID	)			
Tested	09/19/23			
Tester	MHUS			
Recorded	09/19/23			
INSTRUMENTS USED	DURING	INSTRUMENT		VF1-Z28,P124,BID3902,TID411
PREPARATION				23
Sample ID	)			
Tested	09/19/23			
Tester	MHUS			
Recorded	09/19/23			
SHELF LIFE OF PRODUCT		6 WEEKS		6 WEEKS
		TIME		
Sample ID	)			
Tested	09/19/23			
Tester	MHUS			
Recorded	09/19/23			
EXPIRATION DATE		MM/DD/YYYY		10/31/2023
		MM/DD/YYYY		
Sample ID	)			
Tested	09/19/23			
Tester	MHUS			
Recorded	09/19/23			
Made 09/1	9/23			



Page: 2 Date: 09/19/23 at 6:31 PM

## Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results Customer PO # :

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

ichard Stop

**Quality Manager**