

Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results

Customer PO # :

Package	ed Product	DOW023	F		
r ackage		•	a as Nitrogen Standard 1	0 nom Aqueous Solution	
Teet		7111110111			Decult
Test	ED TO FORMU		Target/UOM YES	Range	Result YES
		LATION ON	YES		YES
FILE			YES		
	Sample ID				
	Tested	09/19/23			
	Tester	AROS			
	Recorded	09/19/23	1.00		0.04
PPM NITROGEN 2 DECIMAL		IMAL	1.00	0.90 - 1.10	0.91
PLACES	- ·		PPM		
	Sample ID				
	Tested	09/19/23			
	Tester	AROS			
	Recorded	09/19/23			
APPEARANCE AND COLOR		LOR	CLEAR WATER WHITE		PASS
			APPEARANCE		
	Sample ID				
	Tested	09/19/23			
	Tester	AROS			
	Recorded	09/19/23			
INSTRUMENTS USED DURING		OURING	INSTRUMENT		P154,VF3-J19
PREPARA	ATION				
	Sample ID				
	Tested	09/19/23			
	Tester	AROS			
	Recorded	09/19/23			
EXPIRATION DATE			MM/DD/YYYY		12/30/2023
			MM/DD/YYYY		
	Sample ID				
	Tested	09/19/23			
	Tester	AROS			
	Recorded	09/19/23			
Lo	ot# 30919	970			
	ade 09/19				
	~~~				

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

huland SPA