

Page: 1

Date: 09/21/23 at 4:55 PM

Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569

Certificate of Analysis / QC Results

Customer PO #:

Packaged Product	: 3863			
	FLUORID	E STANDARD 10,000 PPM	as F	
Test		Target/UOM	Range	Result
PREPARED TO FORMU	JLATION ON	YES		YES
FILE		YES		
Sample ID				
Tested	09/21/23			
Tester	MMAT			
Recorded	09/21/23			
PPM FLUORIDE		10000	9990 - 10010	10000
		PPM		
Sample ID				
Tested	09/21/23			
Tester	MMAT			
Recorded	09/21/23			
APPEARANCE AND COLOR		CLEAR LIQUID		PASS
		APPEARANCE		
Sample ID				
Tested	09/21/23			
Tester	MMAT			
Recorded	09/21/23			
INSTRUMENTS USED I		INSTRUMENT		M115,VF4-N21
PREPARATION	20111110			
Sample ID				
Tested	09/21/23			
Tester	MMAT			
Recorded	09/21/23			
EXPIRATION DATE	09/21/23	MM/DD/YY		03/30/24
EXPINATION DATE		MM/DD/YYYY		03/30/24
0		MIM/DD/1111		
Sample ID	00/04/00			
Tested	09/21/23			
Tester	MMAT			
Recorded	09/21/23	OUELE LIEE		CAMONITUO
SHELF LIFE OF PROD	UCT	SHELF LIFE TIME		6 MONTHS
Sample ID				
Tested	09/21/23			
Tester	09/21/23 MMAT			
Recorded	09/21/23			
Lot # 3092166				
Made 09/21/23				
wade 09/2.	1/40			



Page: 2

Date: 09/21/23 at 4:55 PM

Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO #:

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

Quality Manager