

Recorded 09/22/23

Page: 1

Date: 09/22/23 at 1:51 PM

## Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results

## Customer PO # ·

Customer PO #:					
Packaged Product: 8650					
SODIUM	HYDROXIDE 1.0 NORMAL	NIST TRACEABLE SOLUTION			
Test	Target/UOM	Range	Result		
Actual Normality	1.0000	0.9995 - 1.0005	1.0001		
•	EQ/L				
Sample ID					
Tested 09/22/23					
Tester RIRO					
Recorded 09/22/23					
F- Measurement Uncertainty	0.0070	= 0.0070	0.0070		
•	EQ/L				
Sample ID					
Tested 09/22/23					
Tester RIRO					
Recorded 09/22/23					
QCSP7: STANDARDIZATION OF	YES		YES		
AQUEOUS BASES	METHOD				
Sample ID					
Tested 09/22/23					
Tester RIRO					
Recorded 09/22/23					
I.I.S.T. Traceable to SRM 84	Yes		YES		
erri riuosubio to erain e .	YES				
Sample ID	.20				
Tested 09/22/23					
Tester RIRO					
Recorded 09/22/23					
Actual Density	1.04114	1.04049 - 1.04179	1.04107		
ictual Delisity	GM/ML	1.04043 - 1.04173	1.04107		
Sample ID	OW/WIE				
Tested 09/22/23					
Tester RIRO					
Recorded 09/22/23					
PREPARED TO FORMULATION ON	YES		YES		
ILE	YES		ILO		
	ILO				
Sample ID					
Tested 09/22/23					
Tester RIRO Recorded 09/22/23					
	YES		YES		
SAMPLE COLLECTION & HANDLING	YES		IEO		
PER LAB040	IES				
Sample ID					
Tested 09/22/23					
Tester RIRO					
Recorded 09/22/23			00/04/00		
DATE OF SAMPLING FOR TEST	MM/DD/YYYY		09/21/23		
PURPOSES	MM/DD/YYYY				
Sample ID					
Tested 09/22/23					
Tester RIRO					



Page: 2

Date: 09/22/23 at 1:51 PM

## Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results

## Customer PO #:

AMBIENT TEMPERATURE DURING	25.0	10.0 - 35.0	21.7
SAMPLING DEGC	DEG C		
Sample ID			
Tested 09/22/23			
Tester RIRO			
Recorded 09/22/23			
DATE OF TESTING OF SAMPLE	MM/DD/YYYY		09/21/2023
	MM/DD/YYYY		
Sample ID			
Tested 09/22/23			
Tester RIRO			
Recorded 09/22/23			
AMBIENT TEST TEMPERATURE	20.0	15.0 - 30.0	21.7
DEG C	DEG C		
Sample ID			
Tested 09/22/23			
Tester RIRO			
Recorded 09/22/23			
APPEARANCE AND COLOR	CLEAR WATER WHITE		PASS
A I LAWRED AND GOLOR	LIQUID		17.66
	APPEARANCE		
Sample ID	ALLENVANOE		
Tested 09/22/23			
Tester RIRO			
Recorded 09/22/23			
INSTRUMENTS USED DURING	INSTRUMENT		M55,M113,T57
	INSTRUMENT		INI33,INI 13,137
PREPARATION			
Sample ID			
Tested 09/22/23			
Tester RIRO			
Recorded 09/22/23	MM/DD/XX		00/20/05
EXPIRATION DATE	MM/DD/YY		09/30/25
	MM/DD/YYYY		
Sample ID			
Tested 09/22/23			
Tester RIRO			
Recorded 09/22/23	OVEADO		0.724.00
SHELF LIFE OF PRODUCT	2 YEARS		2 YEARS
_	TIME		
Sample ID			
Tested 09/22/23			
Tester RIRO			
Recorded 09/22/23			
<b>Lot #</b> 3091452			
<b>Made</b> 09/22/23			



Page: 3

Date: 09/22/23 at 1:51 PM

Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO #:

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

**Quality Manager**