Page: 1

Date: 09/26/23 at 1:45 PM

Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results

Customer PO #:

Packaged Pro	oduct:	DCP004F			
_		COD STANDARD 100	PPM W/W AQUEOUS		
Test		Targe	t/UOM	Range	Result
PREPARED TO	FORMULATION		ES		YES
FILE		Y	ES		
San	nple ID				
Tes	sted 09/2	26/23			
Tes	ter AR0	OS			
Red	corded 09/2	26/23			
APPEARANCE A	AND COLOR	CLEAR	LIQUID		PASS
		APPEA	RANCE		
San	nple ID				
Tes	sted 09/2	26/23			
Tes	ster AR0	OS			
Red	corded 09/2	26/23			
INSTRUMENTS USED DURING		IG INSTR	UMENT		VF200-T16,VF2-B10
PREPARATION					
San	nple ID				
Tes	sted 09/2	26/23			
Tes	ter AR0	OS			
Red	corded 09/2	26/23			
PPM COD		10	0.0	97.0 - 103.0	100.7
		PI	PM		
San	mple ID				
Tes	sted 09/2	26/23			
Tes	ter AR0	OS			
Red	corded 09/2	26/23			
Lot # 3092011					
Made	09/26/23				

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

Quality Manager

hichard Stol