

Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results

Customer PO # :

Packaged Product:	LY135F			
	MDEA, DEA, NA,	NH3 25.0 PPM SOLUTION		
Test	Targe	et/UOM	Range	Result
PREPARED TO FORMUL		′ES	<u> </u>	YES
FILE	Υ	′ES		
Sample ID				
Tested	09/25/23			
Tester	MMAT			
Recorded	09/25/23			
APPEARANCE AND COL	OR CLEAF	R LIQUID		PASS
	APPE	ARANCE		
Sample ID				
	09/25/23			
Tester	MMAT			
Recorded	09/25/23			
INSTRUMENTS USED DU	URING INSTR	RUMENT		P715,P116,VF1-Z10
PREPARATION				
Sample ID				
Tested	09/25/23			
Tester	MMAT			
Recorded	09/25/23			
EXPIRATION DATE	MM/D	D/YYYY		11/30/2023
	MM/D	D/YYYY		
Sample ID				
Tested	09/25/23			
Tester	MMAT			
Recorded	09/25/23			
Lot # 3092278				
Made 09/25,	/23			

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

hichard For

Quality Manager