



**Aqua Solutions, Inc.**  
**6913 Highway 225**  
**Deer Park, TX, 77536**  
**281/479-2569**  
**Certificate of Analysis / QC Results**  
**Customer PO # :**

<b>Packaged Product:</b> HAC463F			
Chloride Std.Ref2 250mg/L for EZ4000/EZ4006 Series-Chloride			
Test	Target/UOM	Range	Result
<b>PREPARED TO FORMULATION ON FILE</b>	YES YES		YES
Sample ID			
Tested 09/26/23			
Tester AROS			
Recorded 09/26/23			
<b>CALCULATED PPM CHLORIDE</b>	250.00 PPM	248.00 - 252.00	250.00
Sample ID			
Tested 09/26/23			
Tester AROS			
Recorded 09/26/23			
<b>APPEARANCE AND COLOR</b>	CLEAR LIQUID APPEARANCE		PASS
Sample ID			
Tested 09/26/23			
Tester AROS			
Recorded 09/26/23			
<b>INSTRUMENTS USED DURING PREPARATION</b>	INSTRUMENT		M108,VF2-B17
Sample ID			
Tested 09/26/23			
Tester AROS			
Recorded 09/26/23			
<b>EXPIRATION DATE</b>	MM/DD/YYYY MM/DD/YYYY		09/30/2024
Sample ID			
Tested 09/26/23			
Tester AROS			
Recorded 09/26/23			
<b>Lot #</b>	3092635		
<b>Made</b>	09/26/23		

*Richard [Signature]*

**Quality Manager**