

Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results

Customer PO # :

Package	ed Product:				
		0.1% F		Acetonitrile pH 3.42±0.0	
Test			Target/UOM	Range	Result
	D TO FORMUI	LATION ON	YES		YES
FILE			YES		
	Sample ID				
	Tested	09/19/23			
	Tester	JRIO			
	Recorded	09/19/23			
APPEARANCE			CLEAR WATER WHITE		PASS
	Sample ID				
	Tested	09/19/23			
	Tester	JRIO			
	Recorded	09/19/23			
INSTRUMENTS USED DURING		URING	INSTRUMENT		GC,T47
PREPARA	TION				
	Sample ID				
	Tested	09/19/23			
	Tester	JRIO			
	Recorded	09/19/23			
EXPIRATION DATE			MM/DD/YY		03/30/24
			MM/DD/YYYY		
	Sample ID				
	Tested	09/19/23			
	Tester	JRIO			
	Recorded	09/19/23			
PH OF PRODUCT/SOLUTION		TION	3.420	3.380 - 3.460	3.422
			PH		
	Sample ID				
	Tested	09/19/23			
	Tester	JRIO			
	Recorded	09/19/23			
Lo	ot# 30914	28			
Ма	de 09/19	/23			

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

hichard Stor