10/12/23

Made

Page: 1

Date: 10/12/23 at 5:53 PM

Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results

Customer PO #:

Package	d Product:	CG043F	1		
			Solution pH 4.5 ASTM D51	3-16	
Test			Target/UOM	Range	Result
	TO FORMUL	ATION ON	YES		YES
FILE			YES		
	Sample ID				
	Tested	10/12/23			
	Tester	JRIO			
	Recorded	10/12/23			
PH OF PRODUCT/SOLUTION		TION	4.500	4.480 - 4.520	4.502
			PH		
	Sample ID				
	Tested	10/12/23			
	Tester	JRIO			
	Recorded	10/12/23			
APPEARANCE AND COLOR		_OR	CLEAR TO PALE YELLOW		PASS
			LIQUID		
			APPEARANCE		
	Sample ID				
	Tested	10/12/23			
	Tester	JRIO			
	Recorded	10/12/23			
INSTRUME	NTS USED D		INSTRUMENT		M100,GC,EF1-M16
PREPARAT					,,
	Sample ID				
	Tested	10/12/23			
	Tester	JRIO			
	Recorded	10/12/23			
EXPIRATION DATE			MM/DD/YYYY		10/30/2024
			MM/DD/YYYY		
	Sample ID				
	Tested	10/12/23			
	Tester	JRIO			
	Recorded	10/12/23			
	t# 31011				
LOT	[# 51011				

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

1.