

Recorded 10/05/23

Page: 1

Date: 10/05/23 at 10:27 AM

Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results

Customer PO # :

Customer PO #:				
Packaged Product:	1535			
	BUFFER F	REFERENCE pH 13.0 (\pm 0.	.01)	
Test		Target/UOM	Range	Result
PREPARED TO FORMUL	ATION ON	YES		YES
FILE		YES		
Sample ID				
Tested	10/05/23			
Tester	JRIO			
Recorded	10/05/23			
PH OF PRODUCT/SOLUTION		13.000	12.980 - 13.020	12.994
		PH		
Sample ID				
Tested	10/05/23			
Tester	JRIO			
Recorded	10/05/23			
QCSP 20 PH CALIBRATION	ON	YES		YES
METHODS		METHOD		
Sample ID				
Tested	10/05/23			
Tester	JRIO			
	10/05/23			
Actual Temp		25.0	24.9 - 25.1	25.0
		DEG C		
Sample ID				
Tested	10/05/23			
	JRIO			
	10/05/23			
APPEARANCE AND COLOR		CLEAR LIQUID APPEARANCE		PASS
Sample ID				
	10/05/23			
	JRIO			
	10/05/23			
INSTRUMENTS USED DU	IRING	INSTRUMENT		M103,T3,M116
PREPARATION				
Sample ID				
	10/05/23			
	JRIO			
	10/05/23	VEO		V50
pH Traceable to N.I.S.T. S	SKM.2	YES		YES
		YES		
Sample ID	10/05/06			
	10/05/23			
	JRIO 10/05/22			
Recorded PH NIST TRACEABLE TO	10/05/23 SPM 187	YES		YES
FIINISI INACEADLE IU	GRIVI 101	YES		163
Commis ID		ILO		
Sample ID Tested	10/05/22			
	10/05/23			
rester	JRIO			

10/05/23

Made

Page: 2

Date: 10/05/23 at 10:27 AM

Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results

Customer PO #:

PH NIST TRACEABLE TO SRM 2193		YES YES	YES
Sample ID		,	
Tested	10/05/23		
Tester	JRIO		
Recorded	10/05/23		
TRACEABLE TO NIST	SRM 191I AND	YES	YES
SRM 191II		YES	
Sample ID			
Tested	10/05/23		
Tester	JRIO		
Recorded	10/05/23		
EXPIRATION DATE		MM/DD/YY	10/30/24
		MM/DD/YYYY	
Sample ID			
Tested	10/05/23		
Tester	JRIO		
Recorded	10/05/23		
SHELF LIFE OF PRODUCT		1 YEAR	1 YEAR
		TIME	
Sample ID			
Tested	10/05/23		
Tester	JRIO		
Recorded	10/05/23		
Lot # 3090	528		
200 //			

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

Quality Manager

hichard Stol