

Page: 1

Date: 10/13/23 at 11:54 AM

Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569

Certificate of Analysis / QC Results

Customer PO #:

Packaged Product	9577			
	TOC STA	ANDARD 5000 PPM		
Test		Target/UOM	Range	Result
PREPARED TO FORM	ULATION ON	YES		YES
FILE		YES		
Sample ID				
Tested	10/13/23			
Tester	BPHI			
Recorded	10/13/23			
APPEARANCE AND COLOR		CLEAR WATER WHITE		PASS
		LIQUID		
		APPEARANCE		
Sample ID				
Tested	10/13/23			
Tester	BPHI			
Recorded	10/13/23			
INSTRUMENTS USED		INSTRUMENT		M103,T-220
PREPARATION	201	ii to i i toim Ei t i		W.100,1 220
Sample ID				
Tested	10/13/23			
Tester	BPHI			
Recorded				
	10/13/23	MM/DD/YY		04/30/24
EXPIRATION DATE				04/30/24
0 1 10		MM/DD/YYYY		
Sample ID				
Tested	10/13/23			
Tester	BPHI			
Recorded	10/13/23	011515155		0.1401/71/10
SHELF LIFE OF PRODUCT		SHELF LIFE		6 MONTHS
		TIME		
Sample ID				
Tested	10/13/23			
Tester	BPHI			
Recorded	10/13/23			
PPM TOC NO DEC PLA	ACES	5000	4750 - 5250	4960
		PPM		
Sample ID				
Tested	10/13/23			
Tester	BPHI			
Recorded	10/13/23			
Lot # 3101	247			
Made 10/1				
IVIAUE ± 57 ±	-,			



Page: 2

Date: 10/13/23 at 11:54 AM

Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results Customer PO #:

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

Quality Manager