Page: 1

Date: 10/09/23 at 10:27 AM

Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results

Customer PO #:

Packaged Prod	luct: CY369F			
J	SKA 10,	000 PPM NH3 STANDARD		
Test		Target/UOM	Range	Result
PREPARED TO FO	RMULATION ON	YES		YES
FILE		YES		
Samp	e ID			
Tested	10/09/23			
Tester	AROS			
Recor	ded 10/09/23			
CALCULATED PPM NH3		10000	9975 - 10025	10000
		PPM		
Samp	e ID			
Teste				
Tester				
Recor				
INSTRUMENTS USED DURING		INSTRUMENT		M100,DP,VF4-N12
PREPARATION				
Samp	e ID			
Teste				
Tester				
Recor				
EXPIRATION DATE		MM/DD/YYYY		01/30/2024
-XI INATION DATE	=	MM/DD/YYYY		01/00/2024
Samp	o ID			
Samp Tested				
Tester	. 0, 00, 20			
Recorded 10/09/23 APPEARANCE AND COLOR		CLEAR LIQUID		PASS
APPEARANCE AN	COLOR	APPEARANCE		FASS
		APPEARANCE		
Samp				
Tested	. 0, 00, 20			
Tester				
Recor				
Lot # 31	.00923			
Made 10)/09/23			

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

huland St