

Page: 1

Date: 10/06/23 at 1:43 PM

Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results

Customer PO #:

Packaged Produc				
	Fluoric	le Std. 1.0 ppm Ref1 for	EZ3007 Series-Fluoride	
Test		Target/UOM	Range	Result
PREPARED TO FORM	JULATION ON	YES		YES
FILE		YES		
Sample I	D			
Tested	10/06/23			
Tester	JENC			
Recorded	10/06/23			
APPEARANCE AND COLOR		CLEAR WATER WHITE		PASS
		APPEARANCE		
Sample I	D			
Tested	10/06/23			
Tester	JENC			
Recorded	10/06/23			
INSTRUMENTS USED	DURING	INSTRUMENT		P157,VF2-B3
PREPARATION				
Sample I	D			
Tested	10/06/23			
Tester	JENC			
Recorded	10/06/23			
EXPIRATION DATE		MM/DD/YY		04/30/24
		MM/DD/YYYY		
Sample I	D			
Tested	10/06/23			
Tester	JENC			
Recorded	10/06/23			
ACTUAL PPM		1.00	0.90 - 1.10	0.98
		PPM		
Sample I	D			
Tested	10/06/23			
Tester	JENC			
Recorded	10/06/23			
QCSP45:MEASURING	FLUORIDE IN	YES		YES
WATER		QCSP45		
Sample I	D			
Tested	10/06/23			
Tester	JENC			
Recorded	10/06/23			
Lot # 310	0264			
Made 10/06/23				
Made = 07	, = =			



Page: 2

Date: 10/06/23 at 1:43 PM

Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO #:

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

Quality Manager