

Page: 1

Date: 10/13/23 at 2:34 PM

Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results

Customer PO # ·

Customer PO # :					
Packaged Product:	8432				
	SILVER N	NITRATE 0.141 NORMAL			
Test		Target/UOM	Range	Result	
Actual Normality		0.1410	0.1405 - 0.1415	0.1408	
-		EQ/L			
Sample ID					
Tested	10/13/23				
Tester	MHUS				
Recorded	10/13/23				
+- Measurement Uncertainty		0.0022	= 0.0022	0.0022	
		EQ/L			
Sample ID					
Tested	10/13/23				
Tester	MHUS				
Recorded	10/13/23				
SAMPLE COLLECTION	& HANDLING	YES		YES	
PER LAB040		YES			
Sample ID					
Tested	10/13/23				
Tester	MHUS				
Recorded	10/13/23				
QCSP53: SILVER NITRA	ATE STD ON	YES		YES	
AUTO-TITRA		METHOD			
Sample ID					
Tested	10/13/23				
Tester	MHUS				
Recorded	10/13/23				
DATE OF SAMPLING FOR TEST		MM/DD/YYYY		10/13/2023	
PURPOSES		MM/DD/YYYY			
Sample ID					
Tested	10/13/23				
Tester	MHUS				
Recorded	10/13/23				
AMBIENT TEMPERATU	RE DURING	25.0	10.0 - 35.0	22.0	
SAMPLING DEGC		DEG C			
Sample ID					
Tested	10/13/23				
Tester	MHUS				
Recorded	10/13/23				
DATE OF TESTING OF SAMPLE		MM/DD/YYYY MM/DD/YYYY		10/13/2023	
Sample ID					
Tested	10/13/23				
Tester	MHUS				
Recorded	10/13/23				
AMBIENT TEST TEMPERATURE		20.0	15.0 - 30.0	22.0	
DEG C		DEG C			
Sample ID					
Tested	10/13/23				
Tester	MHUS				
Recorded	10/13/23				



Page: 2

Date: 10/13/23 at 2:34 PM

Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results

Customer PO #:

PREPARED TO FORMULATION ON	YES		YES		
FILE	YES				
Sample ID					
Tested 10/13/23					
Tester MHUS					
Recorded 10/13/23					
N.I.S.T. Traceable to SRM 919 NaCl	YES		YES		
	YES				
Sample ID					
Tested 10/13/23					
Tester MHUS					
Recorded 10/13/23					
APPEARANCE AND COLOR	CLEAR WATER WHITE		PASS		
	LIQUID				
	APPEARANCE				
Sample ID					
Tested 10/13/23					
Tester MHUS					
Recorded 10/13/23					
INSTRUMENTS USED DURING	INSTRUMENT		M112,T36,M55,M91,BID4,TID4		
PREPARATION			1123		
Sample ID			5		
Tested 10/13/23					
Tester MHUS					
Recorded 10/13/23	4.04000	4.04700 4.04004	4.04045		
Actual Density	1.01826 GM/ML	1.01792 - 1.01861	1.01815		
0 1 10	GIVI/IVIL				
Sample ID					
Tested 10/13/23					
Tester MHUS					
Recorded 10/13/23					
EXPIRATION DATE	MM/DD/YYYY		04/30/2025		
	MM/DD/YYYY				
Sample ID					
Tested 10/13/23					
Tester MHUS					
Recorded 10/13/23					
Lot # 3100618					
Made 10/13/23					

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Page: 3

Date: 10/13/23 at 2:34 PM

Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO #:

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

Quality Manager