

Page: 1

Date: 10/12/23 at 4:51 PM

Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569

Certificate of Analysis / QC Results
Customer PO #:

Packaged Produc	ct: HAC3471	F		
		Soln for EZ1012/EZ2500	Total Cyanide Analyzer	
Test		Target/UOM	Range	Result
PREPARED TO FORM	JULATION ON	YES		YES
FILE		YES		
Sample I	D			
Tested	10/12/23			
Tester	BPHI			
Recorded	d 10/12/23			
PH OF PRODUCT/SOLUTION		5.200	5.160 - 5.240	5.185
		PH		
Sample I	D			
Tested	10/12/23			
Tester	BPHI			
Recorded				
NORMALITY TO 4 DEC PLACES		0.1003	0.0995 - 0.1011	0.1004
		EQUIV/LITER		
Sample I	D			
Tested	10/12/23			
Tester	BPHI			
Recorded				
APPEARANCE AND C		CLEAR WATER WHITE		PASS
,	,020.1	APPEARANCE		
Sample I	n	711 271171102		
Tested	10/12/23			
Tester	BPHI			
Recorded				
INSTRUMENTS USED		INSTRUMENT		M103,EF6-MAD
PREPARATION		INOTITIONIENT		W100,EI O WAD
	D			
Sample I Tested				
Tester	10/12/23 BPHI			
Recorded				
EXPIRATION DATE	10/12/23	MM/DD/YY		12/17/23
EXPIRATION DATE		MM/DD/YYYY		12/17/23
Sample I	D			
Tested	10/12/23			
Tester	BPHI			
Recorded	d 10/12/23			
Lot # 310	1120			
Made 10/				



Page: 2

Date: 10/12/23 at 4:51 PM

Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO #:

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

Quality Manager