Page: 1

Date: 10/02/23 at 11:08 AM

Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results

Customer PO #:

Package	d Product:	CY369F			
			00 PPM NH3 STANDARD		
Test			Target/UOM	Range	Result
PREPARED TO FORMULATION ON			YES	<u> </u>	YES
FILE			YES		
	Sample ID				
	Tested	10/02/23			
	Tester	SHAR			
	Recorded	10/02/23			
CALCULAT	ED PPM NH3	}	10000	9975 - 10025	10000
			PPM		
	Sample ID				
	Tested	10/02/23			
	Tester	SHAR			
	Recorded	10/02/23			
APPEARANCE AND COLOR			CLEAR LIQUID		PASS
ALL FUIVA	ICL AND CO	-OK	APPEARANCE		1 400
	0		AFFEARANCE		
	Sample ID	10/00/00			
	Tested	10/02/23			
	Tester	SHAR			
	Recorded	10/02/23			
INSTRUMENTS USED DURING		URING	INSTRUMENT		M112,DP,VF4-N17
PREPARA1	ΓΙΟΝ				
	Sample ID				
	Tested	10/02/23			
	Tester	SHAR			
	Recorded	10/02/23			
EXPIRATION DATE			MM/DD/YYYY		01/30/2024
			MM/DD/YYYY		
	Sample ID				
	Tested	10/02/23			
	Tester	SHAR			
	Recorded	10/02/23			
	t# 31002				
Mad	de 10/02	/ 23			

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

huland I