

Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results

Customer PO # :

Packaged Produ	ct: UCT126	ĨF		
	011	STD FOR TSS 40 MG/L SOI	JUTION	
Test	-	Target/UOM	Range	Result
PREPARED TO FOR	MULATION ON	YES	<u>.</u>	YES
FILE		YES		
Sample	ID			
Tested	10/09/23			
Tester	AROS			
Recorde	d 10/09/23			
ACTUAL PPM		40.00	39.50 - 40.50	39.93
		PPM		
Sample	ID			
Tested	10/09/23			
Tester	AROS			
Recorde	d 10/09/23			
APPEARANCE AND COLOR		SLIGHTLY TURBID LIQUID		PASS
		APPEARANCE		
Sample	ID			
Tested	10/09/23			
Tester	AROS			
Recorde	d 10/09/23			
INSTRUMENTS USED DURING		INSTRUMENT		M108,VF4-N17
PREPARATION				
Sample	ID			
Tested	10/09/23			
Tester	AROS			
Recorde	d 10/09/23			
EXPIRATION DATE		MM/DD/YY		04/30/24
		MM/DD/YYYY		
Sample	ID			
Tested	10/09/23			
Tester	AROS			
Recorde	d 10/09/23			
Lot # 310	0912			
	09/23			
made	-			

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

huland SPA