

Page: 1

Date: 10/10/23 at 9:43 AM

Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results

Customer PO #:

Packaged	Product:	TEN222F			
		BROMINE	INDEX STANDARD 18 - 2	0 MG/100G (1 LITER FORMULATION)	
Test			Target/UOM	Range	Result
	TO FORMUL	ATION ON	YES	90	YES
FILE			YES		
	Sample ID				
	Tested	10/10/23			
	Tester	MMAT			
	Recorded	10/10/23			
BROMINE INDEX VALUE			19.00	15.00 - 24.00	20.94
			BROMINE IN		
	Sample ID				
	Tested	10/10/23			
	Tester	MMAT			
	Recorded	10/10/23			
APPEARANCE AND COLOR		.OR	CLEAR WATER WHITE		PASS
			APPEARANCE		
	Sample ID				
	Tested	10/10/23			
	Tester	MMAT			
	Recorded	10/10/23			
INSTRUMEN	NTS USED DI	JRING	LAB EQUIPMENT USED		M100,M115
PREPARATION			INSTRUMENT		
	Sample ID				
	Tested	10/10/23			
	Tester	MMAT			
	Recorded	10/10/23			
EXPIRATION DATE			MM/DD/YY		01/10/24
			MM/DD/YYYY		
	Sample ID				
	Tested	10/10/23			
	Tester	MMAT			
	Recorded	10/10/23			
SHELF LIFE OF PRODUCT		CT	3 MONTHS		3 MONTHS
			TIME		
	Sample ID				
	Tested	10/10/23			
	Tester	MMAT			
	Recorded	10/10/23			
Lot # 3100948					
Made 10/10/23					
ivida	· · · - · ·				



Page: 2

Date: 10/10/23 at 9:43 AM

Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results Customer PO #:

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

Quality Manager