

## Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results

## Customer PO # :

Package	d Product	9594			
. achage		•	NDARD 10,000 PPM (KHP)		
Test			Target/UOM	Range	Result
	D TO FORMU		YES	Rango	YES
FILE			YES		
	Sample ID				
	Tested	10/16/23			
	Tester	DTRA			
	Recorded	10/16/23			
APPEARANCE AND COLOR		LOR	CLEAR LIQUID		PASS
			APPEARANCE		
	Sample ID				
	Tested	10/16/23			
	Tester	DTRA			
	Recorded	10/16/23			
INSTRUMENTS USED DURING			INSTRUMENT		M108,VF12-93
PREPARA	TION				
	Sample ID				
	Tested	10/16/23			
	Tester	DTRA			
	Recorded	10/16/23			
ACTUAL PPM CERTIFIED TOC			10000	9500 - 10500	10170
VALUE			PPM		
	Sample ID				
	Tested	10/16/23			
	Tester	DTRA			
	Recorded	10/16/23			
EXPIRATIO	ON DATE				01/30/24
			MM/DD/YYYY		
	Sample ID	40/40/00			
	Tested	10/16/23			
	Tester Recorded	DTRA			
	Recorded	10/16/23			
Lo	t# 31012				
Ма	<b>de</b> 10/16	5/23			

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

huland SPA